



Published by DiscoverSys

# Evaluation of the accuracy of Bio-impedance in predicting creatinine clearance in ICU patients with acute kidney injury; Comparison with measured creatinine clearance and calculated by formulas



CrossMark

Mohammad Sistanizad,<sup>1,2</sup> Mehran Kouchek,<sup>1</sup> MirMohammad Miri,<sup>1</sup> Sayyed Rasoul Sajjadi,<sup>1</sup> Helena Azimi,<sup>3</sup> Sareh Oushani,<sup>3</sup> Reza Goharani,<sup>1\*</sup>

## ABSTRACT

**Introduction:** It is still of top priority to find a method to measure creatinine clearance (CrCl) in patients with acute kidney injury (AKI) to regulate drug dosage and avoid treatment failure. Therefore, this study designed to compare CrCl calculated with 6-hour urine output, by formulas, and obtained with the bio-impedance method in patients with AKI hospitalized in ICU.

**Materials and Methods:** In this prospective cross-sectional study, 28 adult patients with an increase in serum creatinine of at least 0.3 according to AKIN criteria and admitted in ICU of Imam Hussein Hospital were selected as the study samples. CrCl was calculated by 6-hour urine output collection, through electrical bio-impedance method, and formulas of Cockcroft and Gault, MDRD, and CKD-EPI. After the required data were collected, they were recorded in pre-prepared checklists, and finally fed into software to be statistically analyzed. Lastly, the obtained CrCl values were compared with those measured by 6-hour urine collection to determine the best formula or method to determine CrCl in patients with AKI.

**Results:** The results showed that the bias values of bio-impedance compared to 6-hour urinary clearance, MDRD, Cockcroft, and CKD-EPI formulas were 6.34, 3.31, 5.01, and 15.14, respectively. Also, error values of bio-impedance compared to 6-hour urinary clearance, MDRD, Cockcroft, and CKD-EPI were respectively 140%, 28%, 60%, and 47%. The internal consistency of the bio-impedance values by the above methods were 0.782, 0.983, 0.939, and 0.893, respectively.

**Conclusion:** According to the obtained results, it can be concluded that none of the formulas listed, including MDRD, Cockcroft, and Gault, and CKD-EPI could predict the CrCl correctly compared to 6-hour urine collection as standard. Also, estimated glomerular filtration rate (GFR) by bio-impedance did not reveal acceptable correlation with 6-hour urine collection. Therefore, this group of patients needs a more precise and specific method(s) to determine the CrCl and GFR.

**Keywords:** bio-impedance, creatinine clearance, acute kidney injury (AKI)

**Cite This Article:** Sajjadi, S.R., Kouchek, M., Sistanizad, M., Miri, M.M., Azimi, H., Goharani, R. 2017. Evaluation of the accuracy of Bio-impedance in predicting creatinine clearance in ICU patients with acute kidney injury; Comparison with measured creatinine clearance and calculated by formulas. *Bali Medical Journal* 6(3): 690-695. DOI: [10.15562/bmj.v6i3.758](https://doi.org/10.15562/bmj.v6i3.758)

## INTRODUCTION

Acute kidney injury (AKI) refers to a state in which kidney function decline leads to electrolyte and fluid disorders and acid-base balance following a decrease in GFR. As a common complication in ICU, AKI leads to an increase in mortality among the hospitalized patients<sup>1-4</sup> Almost 7% of patients admitted to hospitals are suffering from AKI.<sup>5</sup> This figure was reported to be 37-67% for patients with specific and critical diseases, depending on different definitions of AKI.<sup>6-9</sup> The incidence of AKI is related to the patients' mortality, such that an approximate mortality rate of 50-70% was reported among patients with AKI who need renal replacement therapy (RRT).<sup>10-13</sup>

Determining serum creatinine (Scr) is the old method of diagnosing AKI in ICU, and etiology of AKI is carried out by evaluating medical background, physical examination, renal ultrasound,

sodium and urea excretion, blood urea nitrogen (BUN) and urine microscopy.<sup>7-9,14</sup> Nowadays, different criteria are utilized to diagnose AKI, including RIFLE and AKIN criteria.<sup>3,8</sup>

One of the major challenges in patients suffering AKI is measurement or prediction of renal function to adjust the dose of medications.<sup>15,16</sup> One of the recommended methods of measuring CrCl is to collect 24-hour urine output and determine clearance through an appropriate formula (Table 1).<sup>17</sup> Due to the problems involved with collecting 24-hour urine output, different studies used shorter periods of time like 2 and 4 hours for critically ill patients to collect urine output and measure CrCl with acceptable correlation with the 24-hour urinary collection.<sup>18-21</sup>

The value of using serum creatinine in estimating GFR in patients with AKI is limited. In fact,

<sup>1</sup>Department of Anesthesiology and Critical Care, Imam Hussein Teaching and Educational Center, Shahid Beheshti University of Medical Sciences, Tehran, Iran

<sup>2</sup>Department of Clinical Pharmacy, Faculty of Pharmacy, Shahid Beheshti University of Medical Sciences, Tehran, Iran

<sup>3</sup>Faculty of Medicine, Shahid Beheshti University of Medical Sciences, Tehran, Iran

\*Correspondence to: Reza Goharani, Department of Anesthesiology and Critical Care, Imam Hussein Teaching and Educational Center, Shahid Beheshti University of Medical Sciences, Tehran, Iran

Received: 2017-07-23

Accepted: 2017-09-7

Published: 2017-09-10

the recommended dosages in most studies and almost in all FDA and EMEA medical brochures are adopted from studies carried out on patients with CKD who have received RRT. In recent studies, it is well shown that estimating CrCl through the proposed formulas such as the Modification of Diet in Renal Disease Study (MDRD), Cockcroft and Gault (CG), and Chronic Kidney Disease Epidemiology Collaboration (CKD-EPI) is not an appropriate criterion for estimating GFR in patients with AKI hospitalized in ICUs, further investigation is required to propose another solution for evaluating renal function in such patients and regulating their drug dosages. Moreover, there is much emphasis on the importance of regulating drug dosage of these patients.<sup>22,23</sup>

BioScan device is an apparatus that operates based on bioelectrical impedance analysis (BIA). Using measured and entered data including height, weight, age, gender, and Scr, BIA presents the CrCl of the patient. Data about the accuracy of this prediction are scanty. The goal of this study was to compare the CrCl measured by 6-hour urine output collection with creatinine clearances calculated using different formulas and bio-impedance method.<sup>22,23</sup>

## METHODS

This prospective cross-sectional study was done in the intensive care unit of a tertiary teaching hospital affiliated to Shahid Beheshti Medical University, Tehran, Iran. In this study, adult ICU patients with an increase in serum creatinine of at least 0.3 g/dL in 48 hours (according to AKIN criteria) were recruited until the desired sample size was obtained. To calculate the sample size, the

maximum value acceptable for intraclass correlation (ICC) of 0.65 and its expected value of 0.85 was taken into consideration. Type 1 error and test power were respectively 0.05 and 0.8. All of the calculations were carried out using ICC.Sample. Size package in R software. Accordingly, the sample size was calculated to be 28. Demographic data of all patients who entered the study were collected and recorded. Moreover, CrCl was calculated by the following methods and relevant formulas: 1) Six-hour urine was collected from all of the patients, the amount of urinary creatinine was measured, and the collected data were used to determine the patients' CrCl as a standard. 2) On the same day of 6-hour urine collection, all of the patients' CrCl and GFR was estimated by electrical bio-impedance method by an intensive care fellowship. 3) CrCl was calculated using the proposed formulas including the simplified refitted MDRD, Cockcroft and Gault, and CKD-EPI for all of the patients.

Lastly, the obtained CrCl was compared with the one measured by 6-hour urine output collection to figure out the best formula or method to determine the CrCl in patients with AKI. The collected data were analyzed using SPSS 22.0. The quantitative data were described through frequency and percentage, and qualitative data through standard deviation, median, and variation range. In the case of non-normal distribution of the data, nonparametric tests were employed. Bland-Altman specialized plots and ICC were utilized to compare the results. The value of  $\alpha$  error was considered as 0.05. In the present study, the required examinations were routine, and no charges were imposed on the patients. Conducting the present study had no contradiction with international treaties in the field of medical science such as Nuremberg and Helsinki codes.

## RESULTS

In the present study, 28 patients were examined. The basic characteristics can be seen in Table 2. The average GFR estimated by different methods can be observed in Table 3.

The bias values of GFR measured through 6-hour urinary clearance compared to bio-impedance, MDRD, CG, and CKD-EPI formulas were 6.34, 9.65, 11.34, and 21.48, respectively. Also, error values of GFR measured through 6-hour urinary clearance compared to bio-impedance, MDRD, CG, and CKD-EPI formulas were 140%, 140%, 163%, and 128% respectively. The internal consistency of GFR values measured through 6-hour urinary clearance by the above methods were 0.782, 0.743, 0.626, and 0.683, respectively.

**Table 1** Formulae for estimating GFR

### Measured 6 hr urinary creatinine clearance

$$\text{CrCl} = (\text{Urine Creatinin} \times \text{daily urine volume}) / (\text{serum Creatinin} \times 1440)$$

### Cockcroft and Gault

$$[(140 - \text{age}) \times \text{weight} / 72 \times \text{sCr}] \times (0.85 \text{ if female})$$

### The simplified refitted MDRD

$$175 \times \text{sCr}^{-1.154} \times \text{age}^{-0.203} (\times 0.742 \text{ if female}) (\times 1.212 \text{ if black})$$

### The CKD-EPI equation

#### • Women:

$$\text{GFR (mL/min/1.73 m}^2) = 144 \times (\text{S-creat}/0.7)^{-0.329} \times 0.993^{\text{age}} (\times 1.15 \text{ if black}) \text{ if S-creat} \leq 0.7 \text{ mg/dL}$$

$$\text{GFR (mL/min/1.73 m}^2) = 144 \times (\text{S-creat}/0.7)^{-1.209} \times 0.993^{\text{age}} (\times 1.15 \text{ if black}) \text{ if S-creat} > 0.7 \text{ mg/dL}$$

#### • Men:

$$\text{GFR (mL/min/1.73 m}^2) = 141 \times (\text{S-creat}/0.9)^{-0.411} \times 0.993^{\text{age}} (\times 1.16 \text{ if black}) \text{ if S-creat} \leq 0.9 \text{ mg/dL}$$

$$\text{GFR (mL/min/1.73 m}^2) = 141 \times (\text{S-creat}/0.9)^{-1.209} \times 0.993^{\text{age}} (\times 1.16 \text{ if black}) \text{ if S-creat} > 0.9 \text{ mg/dL}$$


**Table 2** Baseline characteristics

	Means ± SEM	Min-Max
Age (year)	17.85 ± 60.30	(16-89)
BMI (kg/m <sup>2</sup> )	4.82 ± 26.36	(20-35)
Serum Creatinine(mg/dL)	0.86 ± 1.96	(0.8-4)
Serum Albumin(g/dL)	0.47 ± 2.86	(2.3-4.5)
Serum Urea (mg/dL)	57.02 ± 97.84	(32-263)
Urine Volume(ml/day)	347.19 ± 485.0	(100-1400)
Urine Creatinine (mg/6h)	125.37 ± 167.63	(22-538)
Urine Urea (mg/6h)	7577.76 ± 8136.84	(1000-22000)

**Table 3** GFR estimation

	Min-Max	Means ± SEM
CrCl (ml/min <sup>-1</sup> )	(3.5-119.4)	32.81 ± 32.06
BIA (ml/min <sup>-1</sup> )	(13.6-95.7)	24.17 ± 38.40
MDRD(ml/min <sup>-1</sup> )	(15.5-101.5)	25.97 ± 41.71
CG(ml/min <sup>-1</sup> )	(13.1-114.9)	27.54 ± 43.40
CKD-EPI (ml/min <sup>-1</sup> )	(16.4-128.1)	30.95 ± 53.54

CrCl: 6-hour creatinine clearance , BIA: bioimpedance analyzer , MDRD: The simplified refitted MDRD , CG: Cockcroft and Gault




$$(GFR_{CrCl} + GFR_{BIA}) / 2(\text{ml}/\text{min}^{-1})$$

(Mean : 35.23 ± 26.18 Bias: 6.34 Error(%): 140 Limits of agreement: -41.8/ 54.5)  
(ICC : 0.782 95% CI: 0.549/ 0.896)

**Figure 1** Comparison of GFR measured by CrCl with BIA estimated GFR


The bias values of GFR measured by bio-impedance compared to 6-hour urinary clearance, MDRD, Cockcroft, and CKD-EPI formulas were 6.34, 3.31, 5.01, and 15.14, respectively. Also, error values of GFR measured by bio-impedance compared to 6-hour urinary clearance, MDRD, Cockcroft, and CKD-EPI formulas were 140%, 28%, 60%, and 47% respectively. The internal consistency of GFR measured by bio-impedance values by the



$$(GFR_{CrCl} + GFR_{MDRD}) / 2(\text{ml}/\text{min}^{-1})$$

(Mean : 36.88 ± 26.59 Bias: 9.65 Error(%): 140 Limits of agreement: -61.5/ 42.5)  
(ICC : 0.743 95% CI: 0.463/ 0.877)

**Figure 2** Comparison of GFR measured by CrCl with MDRD calculated GFR



$$(GFR_{CrCl} + GFR_{CKD-EPI}) / 2(\text{ml}/\text{min}^{-1})$$


(Mean : 42.80 ± 28.79 Bias: 21.48 Error(%): 128 Limits of agreement: -76.3/ 33.4)  
(ICC : 0.683 95% CI: 0.145/ 0.867)

**Figure 3** Comparison of GFR measured by CrCl with CKD-EPI calculated GFR

above methods were 0.782, 0.983, 0.939, and 0.893, respectively. These results can be seen in figure 1.


## DISCUSSION

This study showed that none of the formulas, including MDRD, Cockcroft and Gault, and CKD-EPI, could predict the CrCl correctly compared to 6-hour urine collection as standard. Also estimated GFR by bio-impedance did not reveal acceptable correlation with 6-hour urine collection. In spite of routine utilization of the three common formulas of CrCl and 24-hour urine collection in ICU, to the best of our knowledge, the capacity of these methods to precisely evaluate CrCl in critically ill patients with AKI was evaluated only in one study carried out by Bragadottir et al. (2013). They concluded that the minimum difference in calculating and measuring GFR was related to Cr-EDTA, and other methods did not have the appropriate capacity. Moreover, bias in GFR measured by



(Mean :  $37.73 \pm 26.05$  Bias: 11.34 Error(%): 163  
Limits of agreement: -63.4/ 40.7)  
(ICC : 0.626 95% CI: 0.237/ 0.819)

**Figure 4** Comparison of GFR measured by CrCl with CG calculated GFR



Mean :  $40.06 \pm 24.93$  Bias: 3.31 Error(%): 28  
Limits of agreement: -14.5/ 7.8)  
(ICC : 0.981 95% CI: 0.950/ 0.993)

**Figure 5** Comparison of GFR estimating by BIA with MDRD calculated GFR

urinary creatinine is lower than that calculated by using MDRD and Cockcroft-Gault formulas.<sup>24</sup> The main difference between Bragadottir's study and this study is that Bragadottir was using Cr-EDTA as the standard method which has a more precise measuring capacity than 6-hour urine collection. It should be considered that measuring CrCl using 6-hour urine collection is more practical than Cr-EDTA method and is available for all hospitals. Also, it could calculate the CrCl more precise than formulas in patients with AKI. According

to the guidelines of Kidney Disease: Improving Global Outcomes (KIDIGO) in 2012, the formulas proposed to calculate CrCl are validated in patients with CKD but are not specific for those with AKI.<sup>25</sup> Therefore, precise calculation of CrCl in patients with AKI is a dilemma yet. However, because there is no special formula for this group of patients, the standardized formulas for patients with CKD are inevitably utilized.

In 2012, Kirwan et al. studied 51 patients who had AKI based on the AKIN criterion. The results of their study showed that calculating CrCl to measure GFR and evaluate renal function in critically ill patients with AKI is not accurate enough.<sup>23</sup> In another study, Robert et al. compared urinary CrCl with inulin clearance in 20 patients under mechanical ventilation with stable hemodynamics, without inotropic support, and with acute renal dysfunction.<sup>26</sup> The results of their study showed that there was a poor relationship between the two methods of 30-minute and 24-hour creatinine and inulin clearance.

In another study carried out by Erley et al., 24-hour creatinine and inulin clearance in 31 patients with a wide range of renal function hospitalized in ICU were compared.<sup>27</sup> The mean proportion of creatinine to inulin clearance was between 0.54 and 1.92 with a confidence level of 95%, which shows a low level of bias but a high rate of error, i.e., 95% of CrCl values can be 92% higher and 46% lower than inulin clearance values. The results of this study are in line with those of the study conducted by Erley et al.

Another issue relevant to the disagreement among the results of measuring CrCl by different methods is the existence of laboratory errors in both formulas and bio-impedance. However, since all methods have the same basis, i.e., the patient's creatinine, and albumin, this error is likely to run in all methods and cannot be an appropriate reason for justifying the observed difference in measurements.

Within-method repeatability is significant in studies that compare two methods because repeatability of each method can restrict the level of agreement between them.<sup>28,29</sup> Therefore, Bland and Altman's analysis has provided certain criteria to examine the level of agreement between two methods.<sup>28</sup> In an attempt to propose the criteria to examine the level of agreement between two methods, Critchley et al. suggested that for a new method to be accepted, the between-method error should be lower than 30%. They also indicated that intergroup error in each of the both methods and the standard method should be 20% or lower to reach a between-method error of 3%.<sup>30</sup> In this study, the intergroup error was 28% and acceptable only between the two methods of MDRD and BioScan.

The confidence range of 95% was also between 7.8 and 14.5. However, since none of these two methods is used as a standard method for patients with AKI, this finding is not much reliable.

Although 24-hour CrCl is referred to as the standard method, due to its problems in sample collection and waste of time due to prolonged waiting time to see the results, 6-hour CrCl was used instead of 24-hour CrCl in this study. In previous studies in which urinary CrCl was measured using a shorter period of urine output collection from critically ill patients, there was a good relationship among the measured values compared to the longer period.<sup>21,26,31,32</sup>

Estimation formulas are also restricted by serum creatinine as a filtration marker. Precise measurement of GFR based on serum creatinine required a stable state. An increase in serum creatinine level is only observable after a loss of a significant amount of kidney function; therefore, serum creatinine concentration leads to a decrease in GFR and is influenced by other factors that affect renal function. As a result, in unstable conditions, estimating GFR using creatinine-based equations provides incorrect estimations of renal function.<sup>26,33,34</sup>

However, in spite of the significant difference between CrCl by bio-impedance and measured CrCl in this study, further studies should be carried out based on gold standard methods like Cr-EDTA as the reference to obtain more knowledge about this field. The reason for this is the appropriate efficiency of bio-impedance in some studies in other relevant fields.<sup>35-38</sup>

In this study, an appropriate result to estimate GFR was not achieved, it is still of top priority to find a method to measure CrCl in patients with AKI to regulate drug dosage and avoid treatment failure.

## CONCLUSION

According to the results of the present study, it can be stated that neither electrical bio-impedance nor the formulas MDRD, Cockcroft and Gault, and CKD-EPI revealed an acceptable relationship with measuring CrCl by collecting 6-hour urine output as a standard method. Therefore, this group of patients needs a more precise and specific method to measure CrCl and GFR.

## REFERENCES

- Ruiz-Criado J, Ramos-Barron MA, Fernandez-Fresnedo G, Rodrigo E, De Francisco ALM, Arias M, et al. Long-Term Mortality among Hospitalized Non-ICU Patients with Acute Kidney Injury Referred to Nephrology. *Nephron*. 2015;131(1):23-33.
- Bagshaw SM, George C, Bellomo R. Changes in the incidence and outcome for early acute kidney injury in a cohort of Australian intensive care units. *Critical Care*. 2007;11(3):R68.
- Collins AJ, Foley RN, Chavers B, Gilbertson D, Herzog C, Johansen K, et al. US Renal Data System 2011 Annual Data Report. *American Journal of Kidney Diseases*.59(1):A7.
- Waikar SS, Wald R, Chertow GM, Curhan GC, Winkelmayer WC, Liangos O, et al. Validity of International Classification of Diseases, Ninth Revision, Clinical Modification Codes for Acute Renal Failure. *Journal of the American Society of Nephrology*. 2006;17(6):1688-94.
- Nash K, Hafeez A, Hou S. Hospital-acquired renal insufficiency. *American Journal of Kidney Diseases*.39(5):930-6.
- Hoste EAJ, Clermont G, Kersten A, Venkataraman R, Angus DC, De Bacquer D, et al. RIFLE criteria for acute kidney injury are associated with hospital mortality in critically ill patients: a cohort analysis. *Critical Care*. 2006;10(3):R73.
- Mehta RL, Pascual MT, Soroko S, Savage BR, Himmelfarb J, Ikizler TA, et al. Spectrum of acute renal failure in the intensive care unit: The PICARD experience. *Kidney International*.66(4):1613-21.
- Uchino S, Bellomo R, Goldsmith D, Bates S, Ronco C. An assessment of the RIFLE criteria for acute renal failure in hospitalized patients. *Critical care medicine*. 2006;34(7):1913-7.
- Uchino S, Kellum JA, Bellomo R, Doig GS, Morimatsu H, Morgera S, et al. Acute renal failure in critically ill patients: a multinational, multicenter study. *Jama*. 2005;294(7):813-8.
- Jeganathan N, Ahuja N, Yau S, Otu D, Stein B, Balk RA. Impact of End-Stage Renal Disease and Acute Kidney Injury on ICU Outcomes in Patients With Sepsis. *Journal of Intensive Care Medicine*.0(0):0885066616645308.
- Christopher KB. Excess mortality attributable to acute kidney injury in the ICU\*. *Critical care medicine*. 2014;42(4):992-3.
- Bagshaw SM, Laupland KB, Doig CJ, Mortis G, Fick GH, Mucenski M, et al. Prognosis for long-term survival and renal recovery in critically ill patients with severe acute renal failure: a population-based study. *Critical Care*. 2005;9(6):R700-9.
- Liano F, Felipe C, Tenorio MT, Rivera M, Abaira V, Saez-de-Urturi JM, et al. Long-term outcome of acute tubular necrosis: a contribution to its natural history. *Kidney Int*. 2007;71(7):679-86.
- Ostermann M, Chang RW. Acute kidney injury in the intensive care unit according to RIFLE. *Critical care medicine*. 2007;35(8):1837-43; quiz 52.
- Bouchard J, Macedo E, Soroko S, Chertow GM, Himmelfarb J, Ikizler TA, et al. Comparison of methods for estimating glomerular filtration rate in critically ill patients with acute kidney injury. *Nephrology, dialysis, transplantation : official publication of the European Dialysis and Transplant Association - European Renal Association*. 2010;25(1):102-7.
- Jelliffe R. Estimation of creatinine clearance in patients with unstable renal function, without a urine specimen. *American journal of nephrology*. 2002;22(4):320-4.
- Kirwan CJ, Philips BJ, Macphee IA. Estimated glomerular filtration rate correlates poorly with four-hour creatinine clearance in critically ill patients with acute kidney injury. *Critical care research and practice*. 2013;2013:406075.
- Pickering JW, Frampton CM, Walker RJ, Shaw GM, Endre ZH. Four hour creatinine clearance is better than plasma creatinine for monitoring renal function in critically ill patients. *Critical Care*. 2012;16(3):R107.
- Herget-Rosenthal S, Kribben A, Pietruck F, Ross B, Philipp T. Two by two hour creatinine clearance-repeatable and valid. *Clinical nephrology*. 1999;51(6):348-54.
- Cherry RA, Eachempati SR, Hydo L, Barie PS. Accuracy of short-duration creatinine clearance determinations in predicting 24-hour creatinine clearance in critically ill and injured patients. *The Journal of trauma*. 2002;53(2):267-71.
- Herrera-Gutierrez ME, Sellar-Perez G, Banderas-Bravo E, Munoz-Bono J, Lebron-Gallardo M, Fernandez-Ortega JF. Replacement of 24-h creatinine clearance by 2-h creatinine

- clearance in intensive care unit patients: a single-center study. *Intensive care medicine*. 2007;33(11):1900-6.
22. Matzke GR, Aronoff GR, Atkinson AJ, Jr, Bennett WM, Decker BS, Eckardt KU, et al. Drug dosing consideration in patients with acute and chronic kidney disease-a clinical update from Kidney Disease: Improving Global Outcomes (KDIGO). *Kidney Int*. 2011;80(11):1122-37.
  23. Moran SM, Myers BD. Course of acute renal failure studied by a model of creatinine kinetics. *Kidney Int*. 1985;27(6):928-37.
  24. Bragadottir G, Redfors B, Ricksten SE. Assessing glomerular filtration rate (GFR) in critically ill patients with acute kidney injury--true GFR versus urinary creatinine clearance and estimating equations. *Critical Care*. 2013;17(3):R108.
  25. Khwaja A. KDIGO clinical practice guidelines for acute kidney injury. *Nephron Clinical practice*. 2012;120(4):c179-84.
  26. Robert S, Zarowitz BJ, Peterson EL, Dumler F. Predictability of creatinine clearance estimates in critically ill patients. *Critical care medicine*. 1993;21(10):1487-95.
  27. Erley CM, Bader BD, Berger ED, Vochazer A, Jorzik JJ, Dietz K, et al. Plasma clearance of iodine contrast media as a measure of glomerular filtration rate in critically ill patients. *Critical care medicine*. 2001;29(8):1544-50.
  28. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet (London, England)*. 1986;1(8476):307-10.
  29. Mantha S, Roizen MF, Fleisher LA, Thisted R, Foss J. Comparing methods of clinical measurement: reporting standards for bland and altman analysis. *Anesthesia and analgesia*. 2000;90(3):593-602.
  30. Critchley LA, Critchley JA. A meta-analysis of studies using bias and precision statistics to compare cardiac output measurement techniques. *Journal of clinical monitoring and computing*. 1999;15(2):85-91.
  31. Baumann TJ, Staddon JE, Horst HM, Bivins BA. Minimum urine collection periods for accurate determination of creatinine clearance in critically ill patients. *Clinical pharmacology*. 1987;6(5):393-8.
  32. Sladen RN, Endo E, Harrison T. Two-hour versus 22-hour creatinine clearance in critically ill patients. *Anesthesiology*. 1987;67(6):1013-6.
  33. Perrone RD, Madias NE, Levey AS. Serum creatinine as an index of renal function: new insights into old concepts. *Clinical chemistry*. 1992;38(10):1933-53.
  34. Biomarkers of Acute Kidney Injury. *Journal of Toxicology*. 2011;2011.
  35. Thanakitcharu P, Jirajan B. Early detection of subclinical edema in chronic kidney disease patients by bioelectrical impedance analysis. *Journal of the Medical Association of Thailand = Chotmaihet thangphaet*. 2014;97 Suppl 11:S1-10.
  36. Donadio C, Consani C, Ardini M, Caprio F, Grassi G, Lucchesi A. Prediction of glomerular filtration rate from body cell mass and plasma creatinine. *Current drug discovery technologies*. 2004;1(3):221-8.
  37. Peacock WFT. Use of bioimpedance vector analysis in critically ill and cardiorenal patients. *Contributions to nephrology*. 2010;165:226-35.
  38. Maioli M, Toso A, Leoncini M, Musilli N, Bellandi F, Rosner MH, et al. Pre-procedural bioimpedance vectorial analysis of fluid status and prediction of contrast-induced acute kidney injury. *Journal of the American College of Cardiology*. 2014;63(14):1387-94.



This work is licensed under a Creative Commons Attribution